



Section VII. Technical Specifications

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution.]

Procurement of PH Meter Fund 101

ITEM NO.	Specification	UNIT	QTY	BRAND	Statement of Compliance
	PH Meter				
	1. Display: LCD with Green Light Backlighting	unit	2		
	2. Measurement 0-14 Ph, -1999 mV to+1999 mV				
	3. Resolution 0.01 pH, 1 mV				
	4. Accuracy (0.02 pH+2d) (0.5%+ 2d)				
	5. Calibration Automatic buffer Calibration @ pH4, pH7 & pH10				
	6. With Real Time SB Memory Card Data logger (SD card included)				
	7. With automatic temperature Compensation				
	8. With Warranty				
	9. Brand new , Not re-conditioned				
	10. With CE Certification				
	11 Buffer solution pH4 pH7 & pH10				
	Delevery Period 1-2 months after the receipt of purchase order				
	Warranty 1 year				
	For Equipment				
	After Sales Services: Replacement of unit if product is defective				

BIDDER:

