

## Republic of the Philippines ILOILO SCIENCE AND TECHNOLOGY UNIVERSITY

La Paz, Iloilo City
• Trunkline: (033)320-7190 • Telefax: (033)329-4274
• Website: www.isatu.edu.ph



## **Section VII. Technical Specifications**

ITEM NO.		SPECIFICATION	STATEMENT OF COMPLIANCE			
	ABC ( P 250,000.00)		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended salest literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidde or supplier liable for prosecution.]			
ITEM		Procurement of DC POWER SUPP				Statement of
NO.		Specification	UNIT	QTY	BRAND	Compliance
1	DC POWER SUPPLY	60V/3A DC Linear Adjustable Power Supply CV/CC auto switching 3 digits LED display  Line Regulation: CV 1X10^-4+3mV CC 2X10^-3+3mA Load Regulation: CV 1X10^-4+4mV CC 2X10^- 3+5mA  Ripple & Noise: CV 1mVrms CC 3mA rms Display Accuracy: ±(0.5%Rdg+2digits) Input Voltage: 110~127VAC±10%, 220~240VAC±10%	unit	10		
		Delivery Period: 20 Calendar days  Warranty Period: One year warranty on parts and services				