



Republic of the Philippines
ILOILO SCIENCE AND TECHNOLOGY UNIVERSITY

La Paz, Iloilo City

• Trunkline: (033)320-7190 • Telefax: (033)329-4274

• Website: www.isatu.edu.ph



Section VII. Technical Specifications

ITEM NO.	SPECIFICATION	STATEMENT OF COMPLIANCE
	ABC (P 609,000.00)	[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution.]

Procurement of Scientific Laboratory Supplies for CAS Office

	Specification	UNIT	QTY	BRAND	Statement of Compliance
1	FLOW RATE TEMPERATURE Flow rate/Temperature- can measure the rate of movements and temperature, Flow Range 0 m/s to 3.5 m/s Accuracy 0.1 ft/s Pulse Frequency 8.62 pulse/linear foot Unit options meter/sec; feet/sec; total pulses Probe Length 3 to 7 ft. with telescoping tube (Probe is 7 ft when fully expanded) Temperature Range -10°C to 50°C Maximum Length 1.8 m (6 feet) Telescoping handle to reach deep levels Maximum Sample Rate 20 Hz	pieces	5		
2	HEAVY DUTY VOLT METER Heavy duty volt meter Has three ranges, switch-selectable to measure AC or DC voltages. DC Ranges 0 to 3 V/15 V/30 V AC Ranges 0 to 3 V/15 V/30 V Accuracy ±2% Sensitivity 10 kΩ/V	sets	5		
3	METAL PASS TRACK Metal Pass track 1.2m that can fit my pascar ME - 6950 , WITH ACCESSORIES	sets	3		
4	MINI LAUNCHER Mini launcher must have range of 0.5,1,2m launch angle of 0 to 90 degrees and 0 to -45 degrees 1x Launcher Base 1x Loading Rod 1x 2-D Collision Accessory 2x Steel Ball Projectile, 16 mm 1x Safety Glasses 1x Manual 1x Mini Launcher	sets	4		
5	MOTION SENSOR Motion Sensor wireless force sensor Force Range ±50 N Force: Maximum Sampling Rate 1000 Hz Force Resolution 0.03 N Accuracy 0.1 N Acceleration Sensor: Max Sampling Rate 500 Hz Acceleration Range ±16 g Angular Rotation Rate Range up to ±2000 degrees per second Gyro Sensor: Max Sampling Rate 500 Hz Battery Rechargeable lithium-	sets	4		



Republic of the Philippines
ILOILO SCIENCE AND TECHNOLOGY UNIVERSITY

La Paz, Iloilo City

• Trunkline: (033)320-7190 • Telefax: (033)329-4274

• Website: www.isatu.edu.ph



		polymer Logging Yes Bluetooth BT 4.0				
		Delivery Period: 30 days				
		Warranty: 1 year				
